
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002343
10:35 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:54
STERILIZING 3:57
136.0 C 221kPa 3:57
136.4 C 225kPa 4:27
136.3 C 224kPa 4:57
136.4 C 223kPa 5:27
136.3 C 225kPa 5:57
136.2 C 224kPa 6:27
136.3 C 225kPa 6:57
136.3 C 225kPa 7:27
136.3 C 225kPa 7:27
VENTING 7:27
AIR DRYING 8:19
CYCLE COMPLETE 8:29

Digital Signature #
46D732B86D8106FB
