
Midmark M11 - v1.0.5
Total Cycles: 1147

Name: RITTER2
Sterilizer ID: V2313228
05/31/2022 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.8 F 0.0
2:00 186.7 F 3.8
4:00 197.2 F 9.0
6:00 211.6 F 7.8
8:00 226.5 F 9.9
10:00 234.7 F 8.7
12:00 242.1 F 13.3
14:00 242.3 F 12.2
16:00 255.6 F 18.9
18:00 266.7 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.3
1:00 272.4 F 30.0
2:00 271.5 F 29.3
3:00 271.7 F 30.0
4:00 272.4 F 29.9

Min 271.0 F 29.1
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 207.8 F 0.0
10:00 227.1 F 0.0
15:00 222.2 F 0.0
20:00 221.5 F 0.0
25:00 229.7 F 0.0
30:00 228.0 F 0.0

Drying Complete

FILLING: 2:12
HEATING: 19:32
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

05/31/2022 02:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____