
Midmark M11 - v1.0.5
Total Cycles: 4302

Name: RITTER2
Sterilizer ID: V2313228
01/27/2025 07:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.0 F 0.0
2:00 112.0 F 0.9
4:00 114.2 F 3.4
6:00 144.1 F 6.9
8:00 174.8 F 10.8
10:00 202.2 F 4.2
12:00 216.9 F 7.5
14:00 228.3 F 11.8
16:00 230.0 F 8.4
18:00 241.9 F 13.0
20:00 241.5 F 12.2
22:00 253.2 F 18.2
24:00 263.3 F 24.6
26:00 272.3 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 30.0
1:00 272.5 F 30.5
2:00 271.2 F 30.0
3:00 272.3 F 30.5
4:00 271.2 F 29.9

Min 271.1 F 29.8
Max 272.6 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 196.3 F 0.0
10:00 187.9 F 0.0
15:00 181.2 F 0.0
20:00 175.4 F 0.0
25:00 172.5 F 0.0
30:00 173.1 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 26:09
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:05:05
CYCLE COMPLETE

01/27/2025 08:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____