
Midmark M11 - v1.0.5
Total Cycles: 527

Name: RITTER2
Sterilizer ID: V2313228
11/09/2021 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.4 F 0.0
2:00 160.3 F 2.7
4:00 176.4 F 6.8
6:00 202.4 F 12.1
8:00 219.4 F 6.9
10:00 234.1 F 12.5
12:00 235.9 F 10.3
14:00 235.8 F 8.9
16:00 250.6 F 16.2
18:00 262.5 F 23.1
20:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.7 F 29.6
2:00 271.8 F 29.5
3:00 272.5 F 29.7
4:00 271.6 F 29.3

Min 271.6 F 29.1
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 211.5 F 0.0
10:00 219.6 F 0.0
15:00 221.6 F 0.0
20:00 220.8 F 0.0
25:00 221.2 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

11/09/2021 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____