

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4120

Name: RITTER 1  
Sterilizer ID: V2178198  
12/14/2023 04:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.1 F 0.0  
2:00 153.4 F 2.8  
4:00 173.2 F 6.3  
6:00 225.3 F 11.2  
8:00 222.3 F 4.8  
10:00 231.4 F 9.4  
12:00 236.0 F 9.6  
14:00 241.1 F 11.1  
16:00 237.0 F 8.9  
18:00 250.5 F 15.3  
20:00 262.0 F 21.7  
22:00 272.5 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 28.4  
1:00 272.7 F 28.7  
2:00 272.4 F 28.9  
3:00 271.4 F 28.4  
4:00 272.5 F 28.6

Min 271.4 F 28.2  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.6 F 0.0  
5:00 233.5 F 0.0  
10:00 223.9 F 0.0  
15:00 229.1 F 0.0  
20:00 223.6 F 0.0  
25:00 207.6 F 0.0  
30:00 204.6 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 22:58  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:01:43  
CYCLE COMPLETE

12/14/2023 05:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_