

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4255

Name: RITTER2  
Sterilizer ID: V2313228  
01/10/2025 11:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.1 F 0.0  
2:00 170.1 F 0.1  
4:00 179.3 F 1.1  
6:00 199.8 F 5.1  
8:00 219.8 F 9.8  
10:00 231.6 F 8.3  
12:00 237.5 F 10.7  
14:00 234.5 F 8.3  
16:00 241.4 F 10.9  
18:00 247.7 F 14.2  
20:00 259.4 F 20.7  
22:00 269.7 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.6  
1:00 272.5 F 29.0  
2:00 272.2 F 29.0  
3:00 272.0 F 29.2  
4:00 271.9 F 29.2

Min 271.0 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.1 F 0.0  
5:00 204.0 F 0.0  
10:00 189.0 F 0.0  
15:00 188.2 F 0.0  
20:00 187.3 F 0.0  
25:00 187.8 F 0.0  
30:00 187.5 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 22:57  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:01:59  
CYCLE COMPLETE

01/10/2025 12:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_