
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005587
8:32 24/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:45
STERILIZING 3:47
136.3 C 220kPa 3:47
136.1 C 221kPa 4:17
136.2 C 223kPa 4:47
136.1 C 221kPa 5:17
136.4 C 223kPa 5:47
136.3 C 222kPa 6:17
136.4 C 223kPa 6:47
136.2 C 222kPa 7:17
136.2 C 222kPa 7:17
VENTING 7:18
AIR DRYING 8:08
CYCLE COMPLETE 38:08

Digital Signature #
DE77252216772EFB
