
Midmark M11 - v1.0.0
Total Cycles: 4955

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2024 02:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.6 F 0.0
2:00 163.7 F 0.8
4:00 177.2 F 4.1
6:00 200.6 F 8.1
8:00 216.8 F 12.9
10:00 221.8 F 7.0
12:00 234.2 F 11.8
14:00 232.4 F 8.6
16:00 241.7 F 10.9
18:00 244.2 F 12.8
20:00 256.1 F 18.9
22:00 266.7 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.6
1:00 272.5 F 29.2
2:00 272.5 F 29.0
3:00 271.7 F 28.6
4:00 272.9 F 29.4

Min 271.6 F 28.5
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 224.4 F 0.0
10:00 233.5 F 0.0
15:00 227.1 F 0.0
20:00 227.2 F 0.0
25:00 226.6 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 23:30
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:02:22
CYCLE COMPLETE

10/02/2024 03:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____