
Midmark M11 - v1.0.0
Total Cycles: 3399

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2023 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.3 F 0.0
2:00 201.7 F 1.7
4:00 218.0 F 7.1
6:00 231.1 F 12.4
8:00 231.8 F 8.8
10:00 241.1 F 10.5
12:00 243.6 F 12.8
14:00 242.8 F 11.9
16:00 256.1 F 18.7
18:00 267.8 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.3 F 28.9
2:00 271.4 F 28.4
3:00 272.5 F 29.0
4:00 271.6 F 28.4

Min 271.4 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 225.7 F 0.0
10:00 231.1 F 0.0
15:00 229.1 F 0.0
20:00 228.1 F 0.0
25:00 229.6 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 19:18
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

06/15/2023 04:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____