
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001528
12:36 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:24
STERILIZING 3:28
136.4 C 224kPa 3:28
136.4 C 225kPa 3:58
136.2 C 224kPa 4:28
136.4 C 226kPa 4:58
136.4 C 226kPa 5:28
136.4 C 226kPa 5:58
136.2 C 223kPa 6:28
136.4 C 226kPa 6:58
136.4 C 226kPa 6:58
VENTING 6:58
AIR DRYING 7:55
CYCLE COMPLETE 8:00

Digital Signature #
402F351F62F9BDFB
