
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000034
7:39 15/01/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:12
STERILIZING 4:30
136.4 C 221kPa 4:30
136.4 C 224kPa 5:00
136.4 C 223kPa 5:30
136.4 C 222kPa 6:00
136.3 C 222kPa 6:30
136.3 C 222kPa 7:00
136.3 C 223kPa 7:30
136.3 C 224kPa 8:00
136.3 C 224kPa 8:00
VENTING 8:00
AIR DRYING 8:55
CYCLE COMPLETE 9:03

Digital Signature #
9946B0918F165F5B
