

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3391

Name: RITTER 1  
Sterilizer ID: V2178198  
06/14/2023 02:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.0 F 0.0  
2:00 181.8 F 3.3  
4:00 203.1 F 8.7  
6:00 221.4 F 9.1  
8:00 229.7 F 10.0  
10:00 233.3 F 8.3  
12:00 243.5 F 13.3  
14:00 243.2 F 12.2  
16:00 256.3 F 19.0  
18:00 267.5 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.5 F 29.0  
2:00 272.0 F 29.2  
3:00 272.0 F 28.8  
4:00 272.1 F 29.3  
  
Min 271.2 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.4 F 0.0  
5:00 211.0 F 0.0  
10:00 221.2 F 0.0  
15:00 225.5 F 0.0  
20:00 225.5 F 0.0  
25:00 225.0 F 0.0  
30:00 224.3 F 0.0  
  
Drying Complete

FILLING: 1:57  
HEATING: 19:19  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 00:58:44  
CYCLE COMPLETE

06/14/2023 03:55 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_