
Midmark M11 - v1.0.5
Total Cycles: 2464

Name: RITTER2
Sterilizer ID: V2313228
06/05/2023 12:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.4 F 0.0
2:00 159.1 F 3.0
4:00 171.5 F 6.6
6:00 193.3 F 11.6
8:00 217.1 F 6.2
10:00 227.8 F 11.2
12:00 231.6 F 8.0
14:00 242.7 F 12.3
16:00 243.2 F 12.4
18:00 255.5 F 18.9
20:00 266.6 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.5 F 29.7
2:00 271.2 F 29.3
3:00 272.6 F 29.7
4:00 271.1 F 29.1

Min 271.0 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 219.2 F 0.0
10:00 238.4 F 0.0
15:00 225.4 F 0.0
20:00 225.3 F 0.0
25:00 226.2 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:36
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

06/05/2023 01:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____