

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3246

Name: RITTER2  
Sterilizer ID: V2313228  
12/21/2023 03:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 101.7 F 0.0  
2:00 101.3 F 0.2  
4:00 130.4 F 2.1  
6:00 154.2 F 5.4  
8:00 217.3 F 9.5  
10:00 227.5 F 8.6  
12:00 228.0 F 7.5  
14:00 235.3 F 11.9  
16:00 231.8 F 7.9  
18:00 242.8 F 13.2  
20:00 241.5 F 11.1  
22:00 253.3 F 17.2  
24:00 264.3 F 23.6  
26:00 271.8 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.6  
1:00 272.4 F 29.5  
2:00 272.1 F 28.9  
3:00 271.2 F 28.7  
4:00 272.6 F 29.5

Min 271.0 F 28.5  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.8 F 0.0  
5:00 235.0 F 0.0  
10:00 224.2 F 0.0  
15:00 221.5 F 0.0  
20:00 225.5 F 0.0  
25:00 229.4 F 0.0  
30:00 225.3 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 26:10  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:04:20  
CYCLE COMPLETE

12/21/2023 04:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_