

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3910

Name: RITTER 1  
Sterilizer ID: V2178198  
10/17/2023 09:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.4 F 0.0  
2:00 135.7 F 1.3  
4:00 144.1 F 3.7  
6:00 169.3 F 7.1  
8:00 195.7 F 12.1  
10:00 219.1 F 4.8  
12:00 224.6 F 9.3  
14:00 235.3 F 10.5  
16:00 238.2 F 10.4  
18:00 235.3 F 8.1  
20:00 247.9 F 14.1  
22:00 259.5 F 20.5  
24:00 270.3 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.4 F 29.0  
2:00 271.5 F 28.4  
3:00 272.7 F 29.2  
4:00 272.1 F 28.6

Min 271.3 F 28.4  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.1 F 0.0  
5:00 235.6 F 0.0  
10:00 218.6 F 0.0  
15:00 221.5 F 0.0  
20:00 219.9 F 0.0  
25:00 248.8 F 0.0  
30:00 224.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:44  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:04:01  
CYCLE COMPLETE

10/17/2023 10:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_