
Midmark M11 - v1.0.5
Total Cycles: 104

Name: RITTER2
Sterilizer ID: V2313228
06/15/2021 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.3 F 0.0
2:00 110.2 F 1.0
4:00 135.0 F 3.5
6:00 158.7 F 6.5
8:00 182.2 F 10.2
10:00 196.5 F 7.3
12:00 212.3 F 5.4
14:00 224.3 F 9.7
16:00 233.0 F 9.9
18:00 239.5 F 11.7
20:00 239.4 F 10.6
22:00 251.8 F 16.8
24:00 262.8 F 23.3
26:00 272.7 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.8 F 29.7
2:00 271.7 F 29.0
3:00 273.0 F 29.8
4:00 272.0 F 29.1

Min 271.6 F 28.9
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 188.1 F 0.0
10:00 189.0 F 0.0
15:00 188.4 F 0.0
20:00 186.8 F 0.0
25:00 183.7 F 0.0
30:00 193.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 26:20
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:04:43
CYCLE COMPLETE

06/15/2021 04:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____