
Midmark M11 - v1.0.0
Total Cycles: 1865

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2022 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.7 F 0.0
2:00 198.9 F 3.9
4:00 201.4 F 8.7
6:00 228.2 F 8.6
8:00 230.9 F 8.9
10:00 240.7 F 10.3
12:00 241.4 F 11.9
14:00 240.5 F 10.5
16:00 254.5 F 17.5
18:00 266.4 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.2 F 28.6
2:00 271.7 F 28.7
3:00 272.0 F 28.5
4:00 271.9 F 28.8

Min 271.2 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 225.5 F 0.0
10:00 215.2 F 0.0
15:00 214.2 F 0.0
20:00 221.0 F 0.0
25:00 234.4 F 0.0
30:00 227.3 F 0.0

Drying Complete

FILLING: 1:35
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

04/25/2022 04:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____