

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4232

Name: RITTER 1  
Sterilizer ID: V2178198  
01/25/2024 12:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 114.9 F 0.0  
2:00 119.7 F 0.9  
4:00 123.3 F 3.1  
6:00 152.6 F 6.1  
8:00 177.1 F 9.1  
10:00 201.3 F 13.1  
12:00 210.1 F 4.9  
14:00 223.0 F 8.8  
16:00 235.5 F 13.1  
18:00 234.0 F 8.9  
20:00 242.1 F 11.7  
22:00 243.4 F 12.1  
24:00 254.7 F 17.9  
26:00 265.0 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 271.8 F 28.9  
2:00 272.3 F 28.8  
3:00 271.2 F 28.2  
4:00 271.9 F 29.0

Min 271.1 F 28.2  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.3 F 0.0  
5:00 225.4 F 0.0  
10:00 232.6 F 0.0  
15:00 227.8 F 0.0  
20:00 227.3 F 0.0  
25:00 228.9 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 27:52  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:06:34  
CYCLE COMPLETE

01/25/2024 02:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_