
Midmark M11 - v1.0.5
Total Cycles: 191

Name: RITTER2
Sterilizer ID: V2313228
07/12/2021 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.1 F 0.0
2:00 194.1 F 3.7
4:00 209.1 F 9.4
6:00 223.0 F 7.9
8:00 239.9 F 11.4
10:00 239.5 F 10.1
12:00 237.6 F 9.5
14:00 252.9 F 16.8
16:00 264.8 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.1 F 28.6
2:00 272.0 F 29.2
3:00 271.1 F 28.4
4:00 272.1 F 28.9

Min 271.1 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 222.7 F 0.0
10:00 226.7 F 0.0
15:00 228.4 F 0.0
20:00 223.5 F 0.0
25:00 217.7 F 0.0
Dry Cycle Aborted

FILLING: 3:13
HEATING: 17:49
STERILIZING: 4:00
VENTING: 3:08
DRYING: 29:05
TOTAL CYCLE: 00:57:15

07/12/2021 09:33 AM
ACCEPTED: X
REJECTED:
Operator: U01