
Midmark M11 - v1.0.0
Total Cycles: 2925

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2023 02:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.3 F 0.0
2:00 174.9 F 3.9
4:00 192.5 F 8.1
6:00 210.6 F 12.9
8:00 223.6 F 6.6
10:00 235.6 F 11.7
12:00 233.5 F 8.5
14:00 241.5 F 10.6
16:00 245.6 F 13.1
18:00 257.9 F 19.8
20:00 269.1 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.2 F 29.0
2:00 271.2 F 28.6
3:00 272.2 F 29.1
4:00 271.3 F 28.4

Min 271.2 F 28.4
Max 272.3 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 231.4 F 0.0
10:00 232.9 F 0.0
15:00 220.6 F 0.0
20:00 217.1 F 0.0
25:00 223.2 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

02/21/2023 03:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____