
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002172
10:45 28/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:56
STERILIZING 4:00
136.4 C 223kPa 4:00
136.1 C 223kPa 4:30
136.2 C 224kPa 5:00
136.3 C 224kPa 5:30
136.3 C 223kPa 6:00
136.4 C 223kPa 6:30
136.2 C 222kPa 7:00
136.1 C 220kPa 7:30
136.1 C 220kPa 7:30
VENTING 7:31
AIR DRYING 8:22
CYCLE COMPLETE 8:32

Digital Signature #
A89E4A2608E5258E
