

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1966

Name: RITTER2  
Sterilizer ID: V2313228  
01/24/2023 01:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
2:00 187.6 F 3.5  
4:00 203.6 F 8.9  
6:00 220.9 F 8.7  
8:00 236.2 F 10.9  
10:00 234.2 F 8.9  
12:00 238.7 F 9.8  
14:00 248.8 F 15.2  
16:00 261.5 F 22.4  
18:00 272.8 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.4 F 29.4  
2:00 271.8 F 29.3  
3:00 272.5 F 29.4  
4:00 271.4 F 29.0  
  
Min 271.3 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.2 F 0.0  
5:00 233.2 F 0.0  
10:00 221.1 F 0.0  
15:00 208.9 F 0.0  
20:00 211.8 F 0.0  
25:00 222.1 F 0.0  
30:00 219.6 F 0.0  
Drying Complete

FILLING: 1:59  
HEATING: 18:20  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:33  
CYCLE COMPLETE

01/24/2023 02:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_