
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000782
15:05 09/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:35
PRESSURIZING 6:25
STERILIZING 8:09
136.5 C 223kPa 8:09
136.4 C 222kPa 8:39
136.4 C 221kPa 9:09
136.4 C 222kPa 9:39
136.2 C 220kPa 10:09
136.2 C 222kPa 10:39
136.4 C 223kPa 11:09
135.8 C 220kPa 11:39
135.8 C 220kPa 11:39
VENTING 11:39
AIR DRYING 12:41
CYCLE COMPLETE 42:41

Digital Signature #
C98A0FF3492E7EEF
