

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3779

Name: RITTER2  
Sterilizer ID: V2313228  
07/15/2024 08:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.6 F 0.0  
2:00 71.9 F 0.0  
4:00 100.8 F 0.0  
6:00 108.9 F 0.1  
8:00 137.7 F 0.5  
10:00 168.2 F 3.1  
12:00 189.2 F 5.9  
14:00 204.2 F 9.4  
16:00 214.5 F 13.2  
18:00 221.2 F 7.0  
20:00 233.7 F 11.1  
22:00 233.6 F 7.4  
24:00 242.3 F 12.8  
26:00 240.7 F 11.2  
28:00 252.4 F 17.1  
30:00 262.8 F 23.1  
32:00 272.4 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.6 F 29.6  
2:00 271.4 F 29.0  
3:00 272.7 F 29.6  
4:00 271.5 F 28.9  
  
Min 271.3 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 197.6 F 0.0  
10:00 196.2 F 0.0  
15:00 189.9 F 0.0  
20:00 191.0 F 0.0  
25:00 192.2 F 0.0  
30:00 195.3 F 0.0  
  
Drying Complete

FILLING: 0:53  
HEATING: 32:28  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:10:50  
CYCLE COMPLETE

07/15/2024 09:53 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_