
Midmark M11 - v1.0.0
Total Cycles: 5299

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2025 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.6 F 0.0
2:00 181.1 F 2.9
4:00 184.9 F 6.9
6:00 194.2 F 11.1
8:00 208.2 F 3.4
10:00 220.5 F 7.7
12:00 232.4 F 12.3
14:00 234.1 F 8.5
16:00 245.6 F 12.7
18:00 243.5 F 12.0
20:00 254.9 F 17.8
22:00 265.4 F 24.1
24:00 271.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 271.7 F 28.9
2:00 272.3 F 29.2
3:00 272.4 F 29.0
4:00 273.3 F 29.5

Min 271.0 F 28.3
Max 273.3 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 207.4 F 0.0
10:00 211.8 F 0.0
15:00 215.6 F 0.0
20:00 212.5 F 0.0
25:00 215.0 F 0.0
30:00 216.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:06
STERILIZING: 4:00
VENTING: 3:59
DRYING: 30:00
TOTAL CYCLE: 01:03:05
CYCLE COMPLETE

01/30/2025 01:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____