
Midmark M11 - v1.0.0
Total Cycles: 1728

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2022 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 90.3 F 0.0
4:00 125.9 F 0.4
6:00 157.9 F 2.3
8:00 186.9 F 5.8
10:00 207.1 F 10.0
12:00 215.8 F 6.2
14:00 225.6 F 7.3
16:00 237.7 F 12.0
18:00 232.8 F 7.8
20:00 241.6 F 10.8
22:00 244.9 F 12.5
24:00 259.4 F 20.3
26:00 271.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.5 F 28.8
2:00 272.1 F 28.1
3:00 271.9 F 28.3
4:00 272.7 F 28.7

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 224.6 F 0.0
5:00 237.0 F 0.0
10:00 217.7 F 0.0
15:00 215.6 F 0.0
20:00 206.8 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 26:22
STERILIZING: 4:00
VENTING: 2:26
DRYING: 23:45
TOTAL CYCLE: 00:57:40

03/21/2022 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____