
Midmark M11 - v1.0.5
Total Cycles: 976

Name: RITTER2
Sterilizer ID: V2313228
04/11/2022 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.5 F 0.0
2:00 157.5 F 0.1
4:00 174.3 F 1.9
6:00 200.4 F 6.7
8:00 219.0 F 11.9
10:00 227.7 F 8.4
12:00 237.5 F 11.3
14:00 242.7 F 12.9
16:00 243.3 F 12.4
18:00 256.8 F 19.6
20:00 268.6 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.7 F 29.7
2:00 271.6 F 28.9
3:00 272.5 F 29.9
4:00 272.3 F 29.3

Min 271.4 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 213.2 F 0.0
10:00 203.4 F 0.0
15:00 189.4 F 0.0
20:00 188.1 F 0.0
25:00 198.7 F 0.0
30:00 211.8 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

04/12/2022 08:46 AM
ACCEPTED: X
REJECTED:
Operator: U02