
Midmark M11 - v1.0.0
Total Cycles: 5108

Name: RITTER 1
Sterilizer ID: V2178198
11/19/2024 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.6 F 0.0
2:00 156.8 F 0.0
4:00 163.6 F 0.3
6:00 194.7 F 0.6
8:00 213.2 F 3.1
10:00 223.2 F 6.8
12:00 231.1 F 10.8
14:00 233.2 F 7.6
16:00 241.8 F 12.1
18:00 239.0 F 10.0
20:00 237.8 F 8.8
22:00 249.9 F 15.0
24:00 261.1 F 21.3
26:00 271.3 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 271.6 F 28.2
2:00 272.6 F 28.9
3:00 272.4 F 28.6
4:00 271.9 F 28.2

Min 271.4 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 216.8 F 0.0
10:00 221.7 F 0.0
15:00 235.6 F 0.0
20:00 226.7 F 0.0
25:00 225.9 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 26:37
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:10
CYCLE COMPLETE

11/19/2024 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____