
Midmark M11 - v1.0.0
Total Cycles: 4583

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2024 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 74.3 F 0.1
4:00 79.7 F 0.5
6:00 117.2 F 1.3
8:00 147.0 F 3.9
10:00 173.1 F 7.0
12:00 190.9 F 10.5
14:00 199.5 F 4.9
16:00 214.0 F 5.6
18:00 224.8 F 9.1
20:00 235.0 F 13.2
22:00 234.7 F 9.2
24:00 241.1 F 10.8
26:00 244.3 F 12.5
28:00 255.9 F 18.5
30:00 266.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.8 F 29.1
2:00 271.6 F 28.3
3:00 272.8 F 29.2
4:00 272.0 F 28.4

Min 271.5 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 197.2 F 0.0
10:00 198.3 F 0.0
15:00 211.1 F 0.0
20:00 213.1 F 0.0
25:00 214.5 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:44
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:09:50
CYCLE COMPLETE

05/27/2024 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____