
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005328
7:32 10/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:31
STERILIZING 3:42
136.4 C 222kPa 3:42
136.4 C 223kPa 4:12
136.3 C 222kPa 4:42
136.3 C 219kPa 5:12
136.2 C 220kPa 5:42
136.4 C 223kPa 6:12
136.4 C 223kPa 6:42
136.4 C 222kPa 7:12
136.4 C 222kPa 7:12
VENTING 7:12
AIR DRYING 7:58
CYCLE COMPLETE 10:15

Digital Signature #
F45A877196090587
