
Midmark M11 - v1.0.0
Total Cycles: 3441

Name: RITTER 1
Sterilizer ID: V2178198
06/26/2023 07:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 74.9 F 0.0
4:00 82.0 F 0.0
6:00 107.8 F 0.1
8:00 138.1 F 1.4
10:00 166.4 F 3.1
12:00 184.7 F 5.6
14:00 200.2 F 9.0
16:00 213.9 F 12.6
18:00 221.2 F 5.8
20:00 230.9 F 9.8
22:00 237.1 F 10.2
24:00 239.8 F 11.0
26:00 235.3 F 8.5
28:00 249.3 F 14.9
30:00 260.2 F 21.0
32:00 270.2 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 29.1
2:00 271.5 F 28.2
3:00 272.7 F 29.2
4:00 271.9 F 28.5

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 169.8 F 0.0
5:00 202.4 F 0.0
10:00 187.5 F 0.0
15:00 191.3 F 0.0
20:00 214.8 F 0.0
25:00 228.4 F 0.0
30:00 228.5 F 0.0

Drying Complete

FILLING: 0:48
HEATING: 32:57
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:10:56
CYCLE COMPLETE

06/26/2023 08:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____