

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002546  
10:58 28/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 2:02  
STERILIZING 3:09  
136.0 C 222kPa 3:09  
136.0 C 222kPa 3:39  
136.3 C 226kPa 4:09  
136.4 C 226kPa 4:39  
136.4 C 224kPa 5:09  
136.4 C 225kPa 5:39  
136.4 C 224kPa 6:09  
136.4 C 223kPa 6:39  
136.4 C 223kPa 6:39  
VENTING 6:39  
AIR DRYING 7:30  
CYCLE COMPLETE 35:05  
-----

Digital Signature #  
B9C849CF16BF69DC  
-----