
Midmark M11 - v1.0.5
Total Cycles: 670

Name: RITTER2
Sterilizer ID: V2313228
12/30/2021 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.3 F 0.0
2:00 180.0 F 3.1
4:00 197.7 F 8.5
6:00 215.1 F 8.7
8:00 233.2 F 9.5
10:00 236.5 F 9.1
12:00 244.8 F 13.4
14:00 244.6 F 12.7
16:00 257.5 F 19.6
18:00 269.0 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.3 F 29.2
2:00 271.8 F 29.2
3:00 272.1 F 29.0
4:00 271.8 F 29.2

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 246.8 F 0.0
10:00 199.1 F 0.0
15:00 191.1 F 0.0
20:00 209.4 F 0.0
25:00 203.4 F 0.0
30:00 192.3 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

12/30/2021 11:23 AM
ACCEPTED: X
REJECTED:
Operator: U02