
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001159
12:35 10/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:30
STERILIZING 3:27
136.5 C 225kPa 3:27
136.2 C 223kPa 3:57
136.1 C 220kPa 4:27
136.3 C 223kPa 4:57
136.2 C 221kPa 5:27
136.4 C 223kPa 5:57
136.4 C 225kPa 6:27
136.4 C 222kPa 6:57
136.4 C 222kPa 6:57
VENTING 6:57
AIR DRYING 7:51
CYCLE COMPLETE 8:01

Digital Signature #
5BF4EABBC8AD8BB8
