
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000916
18:09 07/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:15
PRESSURIZING 4:16
STERILIZING 5:33
136.0 C 228kPa 5:33
136.2 C 231kPa 6:03
136.3 C 232kPa 6:33
136.4 C 233kPa 7:03
136.4 C 233kPa 7:33
136.3 C 232kPa 8:03
136.1 C 232kPa 8:33
136.4 C 233kPa 9:03
136.4 C 233kPa 9:03
VENTING 9:03
AIR DRYING 10:01
CYCLE COMPLETE 25:01

Digital Signature #
186E7B072D608643
