
Midmark M11 - v1.0.5
Total Cycles: 3366

Name: RITTER2
Sterilizer ID: V2313228
02/07/2024 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.9 F 0.0
2:00 169.8 F 2.6
4:00 187.5 F 6.6
6:00 201.9 F 10.8
8:00 221.8 F 5.6
10:00 230.1 F 9.8
12:00 236.5 F 9.6
14:00 242.5 F 12.9
16:00 242.9 F 12.2
18:00 255.3 F 18.7
20:00 266.5 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.8 F 29.7
2:00 271.8 F 29.0
3:00 271.7 F 29.4
4:00 272.9 F 29.6

Min 271.2 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 218.7 F 0.0
10:00 221.3 F 0.0
15:00 218.6 F 0.0
20:00 217.0 F 0.0
25:00 212.2 F 0.0
30:00 205.8 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:36
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:00:56
CYCLE COMPLETE

02/07/2024 04:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____