

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001830  
7:21 12/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:10  
PRESSURIZING 3:12  
STERILIZING 4:13  
136.4 C 223kPa 4:13  
136.4 C 225kPa 4:43  
136.3 C 223kPa 5:13  
136.2 C 223kPa 5:43  
136.1 C 220kPa 6:13  
136.2 C 221kPa 6:43  
136.2 C 223kPa 7:13  
136.2 C 223kPa 7:43  
136.2 C 223kPa 7:43  
VENTING 7:43  
AIR DRYING 8:29  
CYCLE COMPLETE 9:23  
-----

Digital Signature #  
AE71CCE403EDCFDA  
-----