
Midmark M11 - v1.0.0
Total Cycles: 4628

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2024 09:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.1 F 0.0
2:00 172.4 F 2.5
4:00 177.0 F 6.1
6:00 208.0 F 10.1
8:00 219.4 F 5.4
10:00 226.5 F 8.5
12:00 237.0 F 13.2
14:00 236.2 F 10.0

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 242F
CHAMBER PRESS 12.8PSI
** ITEMS NOT STERILE **

06/12/2024 09:49 AM