

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.2uS / 0.7ppm  
CYCLE NUMBER 001486  
11:12 10/07/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:09  
PRESSURIZING 4:52  
STERILIZING 6:22  
136.4 C 223kPa 6:22  
136.4 C 222kPa 6:52  
136.0 C 219kPa 7:22  
136.5 C 223kPa 7:52  
136.0 C 217kPa 8:22  
136.2 C 221kPa 8:52  
136.2 C 220kPa 9:22  
136.5 C 220kPa 9:52  
VENTING 9:52  
AIR DRYING 10:57  
CYCLE COMPLETE 32:01

-----  
Digital Signature #  
368CC6A1E87D7D36  
-----