
Midmark M11 - v1.0.0
Total Cycles: 3961

Name: RITTER 1
Sterilizer ID: V2178198
10/31/2023 09:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 71.6 F 0.3
4:00 85.6 F 1.7
6:00 124.0 F 4.4
8:00 154.9 F 7.6
10:00 179.2 F 11.5
12:00 204.3 F 2.3
14:00 217.5 F 6.0
16:00 227.6 F 9.7
18:00 234.5 F 10.6
20:00 237.0 F 9.6
22:00 237.9 F 9.3
24:00 245.5 F 12.9
26:00 257.6 F 19.4
28:00 268.4 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.4 F 28.4
2:00 272.6 F 28.8
3:00 272.3 F 28.8
4:00 271.8 F 28.2

Min 271.7 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 226.6 F 0.0
10:00 222.5 F 0.0
15:00 228.3 F 0.0
20:00 228.7 F 0.0
25:00 231.8 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 29:08
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:07:26
CYCLE COMPLETE

10/31/2023 10:33 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____