
Midmark M11 - v1.0.5
Total Cycles: 1807

Name: RITTER2
Sterilizer ID: V2313228
12/05/2022 08:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.1 F 0.0
2:00 176.4 F 3.5
4:00 193.6 F 8.4
6:00 212.4 F 9.5
8:00 228.0 F 8.9
10:00 235.4 F 10.1
12:00 243.8 F 13.2
14:00 244.6 F 12.8
16:00 258.2 F 20.1
18:00 270.3 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.3 F 29.1
2:00 272.4 F 29.6
3:00 272.0 F 28.8
4:00 272.6 F 29.6

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 215.8 F 0.0
10:00 215.1 F 0.0
15:00 215.0 F 0.0
20:00 212.5 F 0.0
25:00 209.9 F 0.0
30:00 207.9 F 0.0

Drying Complete

FILLING: 2:07
HEATING: 18:48
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

12/05/2022 09:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____