
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000388
9:29 25/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:25
CYCLE ABORTED 2:35
CYCLE FAULT 3
NOT STERILE
END TIME 2:35

Digital Signature #
2A8D2890484E48B5
