
Midmark M11 - v1.0.0
Total Cycles: 4520

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2024 08:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 71.1 F 0.0
4:00 80.6 F 0.5
6:00 119.8 F 2.2
8:00 136.3 F 4.8
10:00 205.2 F 8.3
12:00 197.9 F 12.3
14:00 214.3 F 3.3
16:00 217.2 F 6.4
18:00 227.4 F 9.8
20:00 236.5 F 11.2
22:00 236.9 F 9.4
24:00 239.9 F 10.4
26:00 244.5 F 12.2
28:00 255.7 F 18.1
30:00 265.8 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.2 F 28.5
2:00 271.5 F 28.5
3:00 271.7 F 28.3
4:00 272.0 F 28.7

Min 271.1 F 28.0
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 221.8 F 0.0
10:00 218.7 F 0.0
15:00 220.2 F 0.0
20:00 239.7 F 0.0
25:00 228.9 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:50
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:06
CYCLE COMPLETE

05/03/2024 09:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____