
Midmark M11 - v1.0.0
Total Cycles: 818

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2021 04:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.4 F 0.0
2:00 142.7 F 1.6
4:00 167.4 F 3.9
6:00 190.1 F 7.1
8:00 205.8 F 10.9
10:00 220.5 F 5.6
12:00 228.9 F 9.0
14:00 238.7 F 11.9
16:00 239.8 F 11.0
18:00 237.5 F 9.3
20:00 250.4 F 15.4
22:00 261.3 F 21.7
24:00 271.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.3 F 29.1
2:00 272.5 F 28.8
3:00 271.5 F 28.4
4:00 271.3 F 28.4

Min 271.2 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 230.1 F 0.0
10:00 235.6 F 0.0
15:00 220.0 F 0.0
20:00 231.1 F 0.0
25:00 227.0 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 24:44
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:03:38
CYCLE COMPLETE

04/27/2021 05:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____