

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2271

Name: RITTER 1  
Sterilizer ID: V2178198  
08/31/2022 09:16 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 82.7 F 0.0  
2:00 84.4 F 0.0  
4:00 98.4 F 0.1  
6:00 126.3 F 0.3  
8:00 156.1 F 1.0  
10:00 182.1 F 2.2  
12:00 201.9 F 5.5  
14:00 211.9 F 9.2  
16:00 222.9 F 12.8  
18:00 225.4 F 7.0  
20:00 236.9 F 11.2  
22:00 231.9 F 7.3  
24:00 244.3 F 12.9  
26:00 240.7 F 10.7  
28:00 252.0 F 16.3  
30:00 261.8 F 21.9  
32:00 270.9 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.1 F 28.8  
2:00 272.1 F 28.6  
3:00 271.8 F 28.6  
4:00 272.5 F 28.9  
  
Min 271.4 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.8 F 0.0  
5:00 211.9 F 0.0  
10:00 222.4 F 0.0  
15:00 229.4 F 0.0  
20:00 235.9 F 0.0  
25:00 228.8 F 0.0  
30:00 227.2 F 0.0  
  
Drying Complete

FILLING: 0:45  
HEATING: 32:41  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:10:39  
CYCLE COMPLETE

08/31/2022 10:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_