
Midmark M11 - v1.0.0
Total Cycles: 1802

Name: RITTER 1
Sterilizer ID: V2178198
04/07/2022 07:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 74.5 F 0.0
4:00 79.3 F 0.0
6:00 103.9 F 0.1
8:00 141.2 F 1.9
10:00 170.3 F 4.9
12:00 188.5 F 8.6
14:00 210.1 F 12.3
16:00 219.0 F 4.6
18:00 229.2 F 8.5
20:00 239.2 F 12.6
22:00 235.6 F 9.0
24:00 241.2 F 10.6
26:00 245.3 F 12.9
28:00 257.1 F 19.1
30:00 267.8 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.3
1:00 272.6 F 29.0
2:00 271.8 F 28.2
3:00 272.7 F 29.1
4:00 272.1 F 28.3

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 225.8 F 0.0
10:00 224.2 F 0.0
15:00 220.9 F 0.0
20:00 216.7 F 0.0
25:00 223.6 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:14
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:13
CYCLE COMPLETE

04/07/2022 09:02 AM
ACCEPTED: X
REJECTED:
Operator: U02