
Midmark M11 - v1.0.5
Total Cycles: 2729

Name: RITTER2
Sterilizer ID: V2313228
08/03/2023 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.8 F 0.0
2:00 135.5 F 0.0
4:00 158.2 F 0.0
6:00 194.2 F 3.7
8:00 209.7 F 8.0
10:00 222.2 F 12.6
12:00 225.8 F 7.9
14:00 237.9 F 12.9
16:00 238.1 F 10.6
18:00 235.8 F 8.9
20:00 250.0 F 15.8
22:00 261.7 F 22.5
24:00 272.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.7
2:00 271.6 F 29.2
3:00 272.7 F 29.7
4:00 271.7 F 29.0

Min 271.5 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 214.2 F 0.0
10:00 222.0 F 0.0
15:00 207.1 F 0.0
20:00 208.6 F 0.0
25:00 218.2 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:23
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:32
CYCLE COMPLETE

08/03/2023 12:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____