

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3958

Name: RITTER2  
Sterilizer ID: V2313228  
09/19/2024 11:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.1 F 0.0  
2:00 163.1 F 1.7  
4:00 168.8 F 5.3  
6:00 201.2 F 10.1  
8:00 217.0 F 5.2  
10:00 228.7 F 9.6  
12:00 233.5 F 9.5  
14:00 240.8 F 12.3  
16:00 240.1 F 10.9  
18:00 253.0 F 17.5  
20:00 264.7 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 272.7 F 29.6  
2:00 271.1 F 29.1  
3:00 272.2 F 29.8  
4:00 271.5 F 29.0  
  
Min 271.1 F 28.9  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.7 F 0.0  
5:00 197.1 F 0.0  
10:00 197.2 F 0.0  
15:00 193.8 F 0.0  
20:00 194.4 F 0.0  
25:00 194.0 F 0.0  
30:00 192.0 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 21:51  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:00:43  
CYCLE COMPLETE

09/19/2024 12:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_