
Midmark M11 - v1.0.5
Total Cycles: 2432

Name: RITTER2
Sterilizer ID: V2313228
05/30/2023 07:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.4 F 0.0
2:00 75.6 F 0.0
4:00 78.6 F 0.0
6:00 101.9 F 0.0
8:00 146.8 F 2.7
10:00 163.9 F 5.9
12:00 188.7 F 9.5
14:00 212.2 F 13.3
16:00 218.3 F 6.5
18:00 227.9 F 10.4
20:00 233.8 F 8.7
22:00 239.0 F 11.4
24:00 236.4 F 9.1
26:00 248.6 F 15.1
28:00 259.2 F 21.1
30:00 269.0 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.7 F 29.8
2:00 271.5 F 29.0
3:00 272.0 F 29.7
4:00 272.3 F 29.5

Min 271.1 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 198.3 F 0.0
10:00 202.4 F 0.0
15:00 206.2 F 0.0
20:00 208.6 F 0.0
25:00 215.2 F 0.0
30:00 234.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:13
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:12
CYCLE COMPLETE

05/30/2023 08:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____