

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1345

Name: RITTER 1  
Sterilizer ID: V2178198  
11/09/2021 11:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.5 F 0.0  
2:00 133.7 F 0.0  
4:00 176.0 F 0.1  
6:00 189.3 F 2.6  
8:00 204.7 F 6.2  
10:00 221.9 F 10.7  
12:00 229.0 F 5.6  
14:00 234.7 F 10.4  
16:00 237.4 F 7.8  
18:00 245.1 F 13.3  
20:00 243.5 F 11.9  
22:00 256.3 F 18.6  
24:00 267.7 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.4 F 28.9  
2:00 271.5 F 28.1  
3:00 272.4 F 29.0  
4:00 272.0 F 28.4

Min 271.4 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.3 F 0.0  
5:00 211.3 F 0.0  
10:00 221.8 F 0.0  
15:00 226.4 F 0.0  
20:00 228.0 F 0.0  
25:00 230.3 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 25:20  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:03:41  
CYCLE COMPLETE

11/09/2021 12:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_